




PantOs XP

(Space requirements for analog Pan/Ceph, Digital Pan, and Digital Pan/Ceph)

Please, allow a min. of 55in. from center of column to right wall. This will allow minimal space for operator to work.



Each measurement has been rounded up to the nearest inch for minimum requirements.